





**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 1 of 8

**Test Report: EN 1276:2019** 

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional

areas – Test method and requirements (phase 2, step 1)

**Identification of the test laboratory:** Abbott Analytical Ltd

Unit 2, Hickmans Road, Birkenhead, CH41 1JH, Great Britain

**Identification of the client:** IPS Group A/S

Hejreskovvej 22A, DK-3490 Kvistgård, Denmark

**Identification of the sample:** 21G/019

Name of the product: 7100HA

Batch number/reference and

N/A

expiry date (if available):

Date of delivery: 05 July 2021

Storage conditions: Room temperature in darkness

Product diluent recommended by

the manufacturer for use:

Not disclosed

Active substance(s) and their

concentrations (s) (optional):

Not disclosed

Appearance of the product: Blue wipes from which was extracted a clear colourless liquid

#### Notes:

- 1) The test results in this report relate only to the sample(s) tested.
- 2) This test report may not be reproduced except in full, adapted, altered or used to create a derivative work, without written approval from Abbott Analytical Ltd.

Abbott Analytical Ltd Unit 2, Hickmans Road, Birkenhead, CH41 1JH, United Kingdom







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 2 of 8

Test method and its validation:

Method: Dilution-neutralisation

Neutraliser: 100.0 g/l Polysorbate 80 + 30.0 g/l Lecithin +

30.0 g/l Tryptone Soya Broth + 5.0 g/l Sodium thiosulphate +

1.0 g/l L-histidine (Neutraliser B)

Neutraliser validation: Validated in accordance with EN 1276:2019 (5.5.2)

**Experimental conditions:** 

Period of analysis: 21 July 2021 to 23 July 2021

Product test concentration(s): Neat liquid extracted from wipes

Diluent used for product test

solution(s):

Hard water

Contact time(s):  $60 \text{ s} \pm 5 \text{ s}$ Test temperature(s):  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Interfering substance: 0.3 g/l bovine albumin (clean conditions)

Temperature of incubation:  $36^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Identification of the bacterial

Pseudomonas aeruginosa (DSM 939)

strain(s) used: Escherichia coli (NCTC 10418)

Staphylococcus aureus (NCTC 10788) Enterococcus hirae (DSM 3320)

**Deviations:** None

## **Remarks:**

- 1) All test conditions are as requested by the client, irrespective of whether these are in accordance with EN 1276:2019 (5.4.2) or EN 1276:2019 (5.5.1.1).
- 2) Products can only be tested at a concentration of 80% or less as some dilution is always produced by adding the test organisms and interfering substance.







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 3 of 8

# **Requirements:**

The product shall demonstrate at least a 5 decimal log (lg) reduction against every test organism.

#### **Conclusion:**

According to EN 1276:2019, the liquid extracted from this sample of 7100HA wipes possesses bactericidal activity against all of the referenced strains of *Pseudomonas aeruginosa*, *Escherichia coli*, *Staphylococcus aureus* and *Enterococcus hirae*, when tested neat with a contact time of 60 seconds at 20°C under clean conditions.

Approved by:

Signed:

Name: Tony Watson

Position: General Manager

Date: 23 July 2021







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 4 of 8

RST 002 (Issue 5)

Test organism: Pseudomonas aeruginosa (DSM 939)
Date of test: 21 July 2021 Test temperature:  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Interfering substance: 0.3 g/l bovine albumin

Dilution-neutralisation method: Pour plate Number of plates: 1/ml Neutraliser: B Incubation temperature:  $36^{\circ}C \pm 1^{\circ}C$ 

# Validation and controls:

Results: EN 1276:2019

Validation suspension (Nv <sub>0</sub> )			Experimental conditions			Neutraliser or filtration			Method validation (C)		
			control (A)			control (B)			Product conc.: Neat'		Neat*
Vc1	64	<u> </u>	Vc1	65	<u></u> =	Vc1	53	<u> </u>	Vc1	57	<u>n</u> =
Vc2	70	67	Vc2	61	63	Vc2	60	56.5	Vc2	69	63
$30 \le \overline{\varkappa}$ of $Nv_0 \le 160$ ?			$\overline{\mu}$ of $A \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?			$\overline{\mu}$ of $B \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?			$\overline{\mu}$ of $C \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?		
⊠ yes □ no ⊠ yes □ no			⊠ yes	□no		⊠ yes	□no				

Test suspension (N and  $N_0$ ):

Ν	Vc1	Vc2	$\overline{\mu} \text{ wm} = 1.79 \times 10^8  ;$	lg N =	8.25
10 <sup>-6</sup>	188	172	$N_0 = N/10$ ; $\lg N_0 =$	7.25	
10 <sup>-7</sup>	16	18	$7.17 \le \lg N_0 \le 7.70$ ?	⊠ yes	□no

Conc. of the	Contact	Vc1	Vc2	Na	lg <i>Na</i>	lg R
product	time			( $\overline{\mu}$ x 10)		(lg N <sub>o</sub> - lg Na)
Neat*	60 s	0	0	<140	<2.15	>5.10

<sup>\*</sup>Neat liquid extracted from wipes







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 5 of 8

RST 002 (Issue 5)

Test organism: Escherichia coli (NCTC 10418) Date of test: 21 July 2021 Test temperature:  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Interfering substance: 0.3 g/l bovine albumin

Dilution-neutralisation method: Pour plate Number of plates: 1/ml Neutraliser: B Incubation temperature:  $36^{\circ}C \pm 1^{\circ}C$ 

## Validation and controls:

Results: EN 1276:2019

Validation	Validation suspension (Nv <sub>0</sub> )			Experimental conditions			Neutraliser or filtration			Method validation (C)		
			control (A)			control (B)			Product conc.: Neat		Neat*	
Vc1	56	<u> </u>	Vc1	89	<u></u> =	Vc1	57	<u> </u>	Vc1	60	<u>n</u> =	
Vc2	60	58	Vc2	91	90	Vc2	62	59.5	Vc2	60	60	
$30 \le \overline{\mu} \text{ of } Nv_0 \le 160 ? \qquad \overline{\mu} \text{ of } A \ge 0.5$			0.5 x π of Nv₀ ?		$\overline{\mu}$ of $B \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?		ν <sub>o</sub> ?	$\overline{\varkappa}$ of $C \ge 0.5 \times \overline{\varkappa}$ of $Nv_0$ ?		lvo?		
⊠ yes □ no ⊠ yes □ no			⊠ yes	□no		⊠ yes	□no					

Test suspension (N and  $N_0$ ):

Ν	Vc1	Vc2	$\overline{\mu}$ wm = 2.94 x 10 <sup>8</sup> ;	lg N =	8.47
10 <sup>-6</sup>	288	296	$N_0 = N/10$ ; $\lg N_0 =$	7.47	
10 <sup>-7</sup>	34	29	$7.17 \le \lg N_0 \le 7.70$ ?	⊠ yes	□no

Conc. of the	Contact	Vc1	Vc2	Na	lg Na	lg R
product	time			(π x 10)		(lg N <sub>o</sub> - lg Na)
Neat*	60 s	0	0	<140	<2.15	>5.32

<sup>\*</sup>Neat liquid extracted from wipes







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 6 of 8

RST 002 (Issue 5)

Test organism: Staphylococcus aureus (NCTC 10788)

Date of test: 21 July 2021 Test temperature:  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Interfering substance: 0.3 g/l bovine albumin

Dilution-neutralisation method: Pour plate Number of plates: 1/ml Neutraliser: B Incubation temperature:  $36^{\circ}C \pm 1^{\circ}C$ 

#### Validation and controls:

Results: EN 1276:2019

Validation suspension ( $Nv_0$ )			Experimental conditions			Neutraliser or filtration			Method validation (C)		
			control (A)			control (B)			Product conc.: Neat*		Neat*
Vc1	55	<u></u> =	Vc1	54	<u></u> =	Vc1	46	<u> </u>	Vc1	42	<u>n</u> =
Vc2	49	52	Vc2	55	54.5	Vc2	62	54	Vc2	47	44.5
$30 \le \overline{\varkappa}$ of $Nv_0 \le 160$ ?			$\overline{\mu}$ of $A \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?			$\overline{\mu}$ of $B \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?			$\overline{\mu}$ of $C \ge 0.5 \times \overline{\mu}$ of $Nv_0$ ?		
⊠ yes □ no ⊠ yes			⊠ yes	□no		⊠ yes	□no		⊠ yes	□no	

Test suspension (N and  $N_0$ ):

N	Vc1	Vc2	$\pi \text{ wm} = 2.08 \times 10^8 \text{ ;}$	lg N =	8.32
10 -6	224	183	$N_0 = N/10$ ; $\lg N_0 =$	7.32	
10 <sup>-7</sup>	25	26	$7.17 \le \lg N_0 \le 7.70$ ?	⊠ yes	□no

Conc. of the	Contact	Vc1	Vc2	Na	lg Na	lg R
product	time			(π x 10)		(lg N <sub>o</sub> - lg Na)
Neat*	60 s	0	0	<140	<2.15	>5.17

<sup>\*</sup>Neat liquid extracted from wipes







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 7 of 8

RST 002 (Issue 5)

Test organism: Enterococcus hirae (DSM 3320)
Date of test: 21 July 2021 Test temperature:  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ 

Interfering substance: 0.3 g/l bovine albumin

Dilution-neutralisation method: Pour plate Number of plates: 1/ml Neutraliser: B Incubation temperature:  $36^{\circ}C \pm 1^{\circ}C$ 

# Validation and controls:

Results: EN 1276:2019

Validatio	Validation suspension ( $Nv_0$ )			Experimental conditions			Neutraliser or filtration			Method validation (C)		
			control (A)			control (B)			Product conc.: Neat*		Neat*	
Vc1	73	<u> </u>	Vc1	84	<u></u> =	Vc1	71	<u> </u>	Vc1	67	<u>n</u> =	
Vc2	71	72	Vc2	68	76	Vc2	64	67.5	Vc2	63	65	
$30 \le \overline{\varkappa} \text{ of } Nv_0 \le 160 ? \qquad \overline{\varkappa} \text{ of } A \ge 0.5$			.5 x $\overline{\varkappa}$ of Nv <sub>o</sub> ?		$\overline{\varkappa}$ of $B \ge 0.5 \times \overline{\varkappa}$ of $Nv_0$ ?		v <sub>o</sub> ?	$\overline{\varkappa}$ of $C \ge 0.5 \times \overline{\varkappa}$ of $Nv_0$ ?		lvo?		
⊠ yes □ no ⊠ yes □ no			⊠ yes	□no		⊠ yes	□no					

Test suspension (N and  $N_o$ ):

Ν	Vc1	Vc2	$\overline{\mu}$ wm = 3.45 x 10 <sup>8</sup> ;	lg N =	8.54
			$N_0 = N / 10$ ; $\lg N_0 =$		
10 <sup>-7</sup>	34	35	$7.17 \le \lg N_0 \le 7.70$ ?	⊠ yes	□no

Conc. of the	Contact	Vc1	Vc2	Na	lg Na	lg R
product	time			(π x 10)		(lg N <sub>o</sub> - lg Na)
Neat*	60 s	0	0	<140	<2.15	>5.39

<sup>\*</sup>Neat liquid extracted from wipes







**Report:** ADT.21G019.IB **Issued:** 23 July 2021 **Page:** 8 of 8

## **Explanations:**

Vc count per ml (one plate or more)

 $\overline{\mu}$  average of Vc1 and Vc2 (1 + 2 duplicate)

 $\overline{\mu}$  wm weighted mean of  $\overline{\mu}$ 

N number of cells per ml in the test suspension

 $N_0$  number of cells in the test mixture at the beginning of the contact time ( $N_0 = N / 10$ )

Na number of survivors per ml in the test mixture at the end of the contact time (before neutralisation or

filtration)

R reduction ( $\lg R = \lg N_0 - \lg Na$ )

*Nv* number of cells per ml in the validation suspension

 $Nv_0$  number of cells in the validation mixtures at the beginning of the contact time ( $Nv_0 = Nv / 10$ )

A number of survivors per ml in the experimental conditions control mixture

B number of survivors per ml in the neutraliser or filtration control mixture

*C* number of survivors per ml in the method validation mixture

All test results have an associated uncertainty of measurement, details of which are available on request.